

Dartmouth Brain Imaging Center (DBIC)


QA Report 64 CH

20260629


Measurements

Measurements	Value
mean signal (mean)	1938.2
Signal to Noise Ratio (SNR)	215.4
Signal to Fluctuation Ratio (SFNR)	224.8
Percent Fluctuation	0.06
Drift	0.93
Radius of Decorrelation (RDC)	7.3
Mean Ghost Percentage	2.139
Standard Deviation (std)	1.13


Signal




Frequency Spectrum




Radius of Decorrelation




Smoothness - X




Smoothness - Y




Smoothness - Z




Center of Mass - X




Center of Mass - Y




Center of Mass - Z




Ghost




Odd-Even Difference Image



Mean Image



Standard Deviation Image



SFNR Image

